

Patient: CONRADO ARGUELLO PERALTA 59A (#32532532)
Series: Dose Record (#997)
Manufacturer: GE Medical Systems (Optima CT520 Series, #*)
Completion Flag: COMPLETE
Verification Flag: UNVERIFIED
Content Date/Time: 2026-08-28 18:37:46

X-Ray Radiation Dose Report

(observed: 2026-08-28 18:39:17)

Concept Modifier: Procedure reported [[Annex 1](#)]
Observation Context: Observer Type = Device (121007, DCM)
Observation Context: Device Observer UID = 1.2.840.113619.6.278
Observation Context: Device Observer Name = "FOUCT"
Observation Context: Device Observer Manufacturer = "GE Medical Systems"
Observation Context: Device Observer Model Name = "Optima CT520 Series"
Observation Context: Device Observer Serial Number = "*"
Observation Context: Device Role in Procedure = Irradiating Device (113859, DCM)
Observation Context: Device Role in Procedure = Recording (121097, DCM)
Observation Context: Start of X-Ray Irradiation = 2026-08-28 18:37:46
Observation Context: End of X-Ray Irradiation = 2026-08-28 18:39:17
Observation Context: Scope of Accumulation [[Annex 2](#)]

CT Accumulated Dose Data

2 {events} 2418.56 mGy.cm

CT Acquisition

1.1 CRANEO SIMPLE TCE Head Constant Angle Acquisition
1.2.840.113619.2.278.3.2831196616.183.1787958580.894

CT Acquisition Parameters

3.0 s 308.36 mm 154.36 mm -154.0 mm 1.2.840.113619.2.278.3.2831196616.183.1787958580.892.5655.1
1.25 mm 308.36 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 10.0 mA 10.0 mA

CT Acquisition

1.1 CRANEO SIMPLE TCE Head Spiral Acquisition
1.2.840.113619.2.278.3.2831196616.183.1787958580.897

CT Acquisition Parameters

20.01 s 225.08 mm 201.75 mm 246.1 mm 61.54 mm -140.21 mm 73.15 mm -151.93 mm
1.2.840.113619.2.278.3.2831196616.183.1787958580.892.5655.1 1.25 mm 20.0 mm 0.56 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 350.0 mA 350.0 mA 1.0 s

CT Dose

107.45 mGy IEC Head Dosimetry Phantom 2418.56 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

Source of Dose Information:

Automated Data Collection (113856, DCM)

Annex

Annex 1

Procedure reported:

Computed Tomography X-Ray (P5-08000, SRT)

Concept Modifier: Has Intent = Diagnostic Intent (R-408C3, SRT)

Annex 2

Scope of Accumulation:

Study (113014, DCM)

Properties: Study Instance UID = 1.2.840.113619.2.278.3.2831196616.183.1787958580.890